

Issued by NMI Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

In accordance with The Council Directive 2009/23/EC on non-automatic weighing instruments.

Manufacturer SPES Bilance S.n.c..
Via Bonacalza 3
21040 Oggiona Santo Stefano (VA)
Italy

In respect of A class **(III)**, electronic, single- or multi-interval **non-automatic weighing instrument**.

Manufacturers name : SPES BOSCHIROLI FAUSTINO DI BOSCHIROLI
MANUELA E CLAUDIO & C. SNC

Type : JP / JSP / AHR (direct sales to the public)
JW / JWP (not for direct sales to the public)

Characteristics $3 \text{ kg} \leq \text{Max} \leq 45 \text{ kg}$
 $e \geq 1 \text{ g}$
 $n \leq 3000$ divisions (per partial weighing range)
Maximum of 2 partial weighing ranges
 $T \leq -\text{Max}$
Temperature range $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$

In the description number T7953 revision 0 further characteristics are described.

Valid until 14 November 2021

Description and documentation The instrument is described in the description number T7953 revision 0 and documented in the documentation folder T7953-1, appertaining to this EC type-approval certificate.

The Notified Body no. 0122
NMI Certin, 14 November 2011



C. Oosterman
Head Certification Board