

Issued by NMI Certin B.V.  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

In accordance with The Council Directive 2009/23/EC on non-automatic weighing instruments.

Manufacturer SPES Bilance S.n.c.  
Via Bonacalza 3  
21040  
Oggiona Santo Stefano (VA)  
Italy

In respect of A class **(III)**, electronic, single- or multi-interval, multi-range **non-automatic weighing instrument**, intended to be used for direct sales to the public.  
Manufacturers name : SPES BOSCHIROLI FAUSTINO DI BOSCHIROLI  
MANUELA E CLAUDIO & C. SNC  
Type : MR10

Characteristics  $3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$   
 $e \geq 1 \text{ g}$   
 $n \leq 6000$  divisions for single interval and multi-range instruments  
 $n \leq 3000$  divisions per partial weighing range for multi-interval instruments  
Maximum of 2 partial weighing ranges  
Temperature range -10 °C / +40 °C  
In the description number T7954 revision 0 further characteristics are described.

Valid until 14 November 2021

Description and documentation The instrument is described in the description number T7954 revision 0 and documented in the documentation folder T7954-1, appertaining to this EC type-approval certificate.

The Notified Body no. 0122  
NMI Certin, 14 November 2011

  
C. Oosterman  
Head Certification Board

**NMI Certin B.V.**  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMI Certin BV. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI (see "Regulation objection and appeal against decisions of NMI" [www.nmi.nl](http://www.nmi.nl))

Reproduction of the complete document only is permitted