

Issued by	NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands
In accordance with	The Council Directive 2009/23/EC on non-automatic weighing instruments.
Manufacturer	SPES Bilance S.n.c. Via Bonacalza 3 21040 Oggiona Santo Stefano (VA) Italy
In respect of	A class <b>(III)</b> , electronic, single- or multi-interval, <b>non-automatic weighing instrument</b> , intended to be used for direct sales to the public. Manufacturers name : SPES BOSCHIROLI FAUSTINO DI BOSCHIROLI MANUELA E CLAUDIO & C. SNC Type : RIO
Characteristics	$6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$ $e \geq 1 \text{ g}$ $n \leq 6000$ divisions (in case of single interval) $n \leq 3000$ divisions per partial weighing range (in case of multi interval) Maximum of 2 partial weighing ranges $T \leq - \text{Max}$ Temperature range $0 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$  In the description number T7955 revision 0 further characteristics are described.
Valid until	14 November 2021
Description and documentation	The instrument is described in the description number T7955 revision 0 and documented in the documentation folder T7955-1, appertaining to this EC type-approval certificate.

The Notified Body no. 0122  
NMi Certin, 14 November 2011



C. Oosterman  
Head Certification Board

**NMi Certin B.V.**  
Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
T +31 78 6332332  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi" [www.nmi.nl](http://www.nmi.nl))

Reproduction of the complete document only is permitted