

Issued by	NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands
In accordance with	The Council Directive 2009/23/EC on non-automatic weighing instruments.
Manufacturer	Dikomsan Elektronik Sanayi Ve Ticaret Ltd. Sti. Menderes Cad. 19 4. Levent 34418 Istanbul Turkey
In respect of	A class (III) , electronic, single- or multi-interval, non-automatic weighing instrument , intended to be used for direct sales to the public. Manufacturer mark/name: Dikomsan Type : ML Series (ML-N, ML-P, ML-G)
Characteristics	$6 \text{ kg} \leq \text{Max} \leq 32 \text{ kg}$ $e \geq 1 \text{ g}$ $n \leq 6400$ divisions in case of single-interval $n \leq 3000$ divisions in case of multi-interval Maximum of 2 partial weighing ranges $T \leq -50\%$ of Max Temperature range $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$ In the description number T8085 revision 1 further characteristics are described.
Valid until	5 March 2022
Description and documentation	The instrument is described in the description number T8085 revision 1 and documented in the documentation folder T8085-1, appertaining to this EC type-approval certificate.
Remark	This revision EC type-approval certificate replaces the earlier version, except for its documentation folder.

The Notified Body no. 0122
NMi Certin, 23 March 2012



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi" www.nmi.nl)

Reproduction of the complete document only is permitted