

Issued by	NMi Certin B.V. Hugo de Grootplein 1 3314 EG Dordrecht The Netherlands
In accordance with	The Council Directive 2009/23/EC on non-automatic weighing instruments.
Manufacturer	DIGI France S.A. Z.A. Central Parc 4 Allée du Sanglier 93421 Villepinte France
In respect of	A class (III) , electronic, single- or multi-interval non-automatic weighing instrument , intended to be used for direct sales to the public (only for DS-673). Manufacturer mark/name : DIGI France S.A. Type : DS-673(SS)
Characteristics	$1,5 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$ $e \geq 0,5 \text{ g}$ $n \leq 7500$ divisions (for single interval instruments) $n \leq 3000$ divisions for multi interval instruments (per partial weighing range) Maximum of 2 partial weighing ranges Temperature range $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$ for DS-673 and $0 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$ for DS-673SS In the description number T8214 revision 1 further characteristics are described.
Valid until	10 July 2022
Description and documentation	The instrument is described in the description number T8214 revision 1 and documented in the documentation folder T8214-2, appertaining to this EC type-approval certificate.
Remark	This EC type-approval certificate replaces the earlier version, including its documentation folder.
The Notified Body no. 0122	NMi Certin, 10 August 2012

C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The designation of NMi Certin BV. as Notified Body can be verified at <http://ec.europa.eu/enterprise/newapproach/nando/>

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see "Regulation objection and appeal against decisions of NMi" www.nmi.nl)

Reproduction of the complete document only is permitted