

Issued by	NMi Certin B.V., designated and notified by the Netherlands to perform tasks with respect to conformity assessment procedures mentioned in Article 13 of Directive 2014/31/EU, after having established that the measuring instrument meets the applicable requirements of Directive 2014/31/EU, to:
Manufacturer	DIGI Europe Ltd. DIGI House, Rookwood Way Haverhill Suffolk, CB9 8DG United Kingdom
Measuring instrument	A Non-automatic weighing instrument Type : DPS-800s, DPS-800 II Further properties are described in the annex: – Description T12179 revision 1.
Initially issued	12 May 2021
Valid until	12 May 2031
Remark	This revision replaces the earlier version.

Issuing Authority **NMi Certin B.V., Notified Body number 0122**
21 March 2024

Certification Board

NMi Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the manufacturer shall indemnify third-party liability.

The designation of NMi Certin B.V. as Notified Body can be verified at <http://ec.europa.eu/growth/tools-databases/nando/>

Reproduction of the complete document only is permitted.

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon on top of the electronic version of this certificate.

1 General information about the non-automatic weighing instrument

All properties of the non-automatic weighing instrument, whether mentioned or not, shall not be in conflict with the legislation.

This certificate contains references to other certificates. The properties mentioned in these certificates shall be observed in addition to the properties mentioned in this certificate.

1.1 Essential parts

Indicator:

Producer	Type	Certificate number
DIGI Europe Ltd.	DPS-800s, DPS-800 II	TC11877

Any analog passive load cell(s) may be used under this certificate for instruments as described in WELMEC 2.4 Issue 2021, provided the following conditions are met:

- There is a respective certificate (EN45501) or an OIML Certificate (R 60) issued for the load cell by a Notified Body responsible for type examination under Directive 2014/31/EU or Directive 2014/32/EU;
- The load cell shall comply with humidity class **CH** or **SH**;
- The certificate contains the load cell types and the necessary load cell data required for the manufacturer's declaration of compatibility of modules (EN 45501:2015 clause F.4), and any particular installation requirements;
- The compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in EN 45501:2015 clause F.4, at the time of putting into use;
- The load transmission must conform to one of the examples shown in WELMEC 2.4 Issue 2021.

1.2 Essential characteristics

Accuracy class	Ⓜ or ⓂⓂ
Maximum number of scale intervals	$n \leq$ number of scale intervals in the certificates involved

1.3 Essential shapes

The data plate is secured against removal by sealing or will be destroyed when removed.

1.4 Conditional parts

The instrument may be equipped with the following parts that further process the measurement result without modification under the conditions stated in the table:

Part	Condition(s)	Reference document
Simple recipient printer	CE marking present	WELMEC 2.10 clause 3.1.3
Printer Data Storage Device	CE marking present and the part is certified to be connected to a weighing instrument by a Notified Body responsible for type examination under Directive 2014/31/EU.	WELMEC 2.10 clause 3.1.3

1.5 Non-essential parts

The non-automatic weighing instrument may be connected to non-essential devices, for example but not limited to bar code readers, foot switches, second displays and cash drawers, provided that:

- They do not present primary data used for purposes mentioned in Directive 2014/31/EU Article 1(2), (a) to (f) unless the "Preliminary observation" in Directive 2014/31/EU Annex I is satisfied;
- They do not lead to an instrument having other essential characteristics than those fixed by this certificate.

2 Seals

To secure components that may not be dismantled or adjusted by the user, the non-automatic weighing instrument has to be secured in a suitable manner on the locations indicated in the drawing and certificates involved.

The connecting cable of the load cell or the junction box is provided with the possibility to seal.

3 Conditions for conformity assessment

The marks, facilities for the marks and the inscriptions on the non-automatic weighing instrument fulfil the requirements of Directive 2014/31/EU Annex III clause 1.

The compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in EN 45501:2015 clause F.4, at the time of putting into use.