

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity procedures mentioned in Article 13 of Directive 2014/31/EU,
after having established that the measuring instrument meets the
applicable requirements of Directive 2014/31/EU, to:

Manufacturer Ebinger Waagenbau GmbH
Obere Riedwiesen 15
74427 Fichtenberg
Germany

Measuring instrument **A Non-automatic weighing instrument**
Type : EBi-D33XW, EBi -D33P, EBD52P, EBD52XW, EBi-D61XW,
EBi-D61PW, EB72XW, EB82XWT

Further properties are described in the annex:
– Description T12417 revision 0.

Valid until 28 June 2032

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
28 June 2022

Certification Board

NMI Certin B.V.
Thijsseweg 11
2629 JA Delft
The Netherlands
T +31 88 6362332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision
that no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMI Certin B.V. as Notified
Body can be verified at
[http://ec.europa.eu/growth/tools-
databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete
document only is permitted.

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon
on top of the electronic version of
this certificate.

1 General information about the non-automatic weighing instrument

All properties of the non-automatic weighing instrument, whether mentioned or not, shall not be in conflict with the legislation.

This certificate contains references to other certificates. The properties mentioned in these certificates shall be observed in addition to the properties mentioned in this certificate.

1.1 Essential parts

Indicator:

Producer	Type	Certificate number
Ohaus Corporation	i-DT33XW, i -DT33P	TC12208
Ohaus Corporation	TD52P, TD52XW	TC11214
Ohaus Corporation	i-DT61XW, i-DT61PW	TC11888
Ohaus Corporation	T72XW	TC8509
Ohaus Corporation	T82XWT	0200-WL-07557

Any analog passive load cell(s) may be used under this certificate for instruments as described in WELMEC 2.4 Issue 2021, provided the following conditions are met:

- There is a respective certificate (EN45501) or an OIML Certificate (R 60) issued for the load cell by a Notified Body responsible for type examination under Directive 2014/31/EU or Directive 2014/32/EU.
- The load cell shall comply with humidity class **CH** or **SH**;
- The certificate contains the load cell types and the necessary load cell data required for the manufacturer's declaration of compatibility of modules (EN 45501:2015 clause F.4), and any particular installation requirements;
- The compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in EN 45501:2015 clause F.4, at the time of putting into use;
- The load transmission must conform to one of the examples shown in WELMEC 2.4 Issue 2021.

1.2 Essential shapes

The data plate is secured against removal by sealing or will be destroyed when removed.

1.3 Conditional parts

The instrument may be equipped with the following parts that further process the measurement result without modification under the conditions stated in the table:

Part	Condition(s)	Reference document
Simple recipient printer	CE marking present	WELMEC 2.10 clause 3.1.3
Printer Data Storage Device	CE marking present and the part is certified to be connected to a weighing instrument by a Notified Body responsible for type examination under Directive 2014/31/EU.	WELMEC 2.10 clause 3.1.3

The non-automatic weighing instrument is fitted with a levelling device and a level indicator, unless the instrument is installed in a fixed position.

1.4 Non-essential parts

The non-automatic weighing instrument may be connected to non-essential devices, for example but not limited to bar code readers, foot switches, second displays and cash drawers, provided that:

- They do not present primary data used for purposes mentioned in Directive 2014/31/EU Article 1(2), (a) to (f) unless the (Preliminary observation) in Directive 2014/31/EU Annex I is satisfied;
- They do not lead to an instrument having other essential characteristics than those fixed by this certificate.

2 Seals

To secure components that may not be dismantled or adjusted by the user, the non-automatic weighing instrument has to be secured in a suitable manner on the locations indicated in the certificates involved.

3 Conditions for conformity assessment

The marks, facilities for the marks and the inscriptions on the non-automatic weighing instrument fulfil the requirements of Directive 2014/31/EU Annex III clause 1.

The compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in EN 45501:2015 clause F.4 at the time of putting into use.